


# Inspection sheet


Date 2001/4/11

Model FT/IR-460  
Serial No. A004060846  
Temperature 14 C  
Humidity Unk %  
Operator JASCO (UK) Ltd  
Total Judgement good

Inspector



Inspection :%T Change



[Standard] Standard +/-5 %T

Wavenumber	Measured	Standard	Difference	Judgement
4000.00	90.866	90.683	0.184	good
3000.00	66.298	66.193	0.105	good