


Inspection sheet


Date 2001/4/11

Model FT/IR-460
Serial No. A004060846
Temperature 14 C
Humidity Unk %
Operator JASCO (UK) Ltd
Total Judgement good

Inspector



Inspection :Background



[Standard] Acceptance criterion 60 %

Wavenumber	Measured	Standard	Meas / Std	Judgement
3000.00	70.769	70.771	99.998 %	good
2000.00	127.353	127.665	99.755 %	good
1600.00	111.478	111.226	100.227 %	good
1000.00	41.539	41.538	100.003 %	good