


# Inspection sheet


Date 2001/4/11

Model FT/IR-460  
Serial No. A004060846  
Temperature 14 C  
Humidity Unk %  
Operator JASCO (UK) Ltd  
Total Judgement good

Inspector



Inspection :WN Accuracy



[Standard] Acceptance criterion +/-2 cm-1

STD WN	Measured	Difference	Judgement
1601.40	1601.110	-0.290	good
906.70	906.862	0.162	good