


# Inspection sheet


Date 2001/4/11

Model FT/IR-460  
Serial No. A004060846  
Temperature 14 C  
Humidity Unk %  
Operator JASCO (UK) Ltd  
Total Judgement good

Inspector



Inspection :WN Repeatability



[Standard] +/-1 cm-1 (4600 - 400cm-1)

Peak[cm-1]	1st	2nd	Difference	Judgement
2850.70	2848.828	2848.828	0.000	good
1601.40	1601.110	1601.110	0.000	good
1583.10	1582.789	1582.789	0.000	good
1181.40	1181.186	1181.186	0.000	good
1154.30	1154.670	1154.670	0.000	good
1069.10	1069.335	1069.335	0.000	good
1028.00	1028.355	1028.355	0.000	good
906.70	906.862	906.862	0.000	good